

TECHNICAL INSERT

Item No	575	ChemDisk™ Monitor for	Prepared by	SCW
Edition	11/18/2009	Nitrous Oxide	Approved by	CRM

The ChemDisk™ Personal Monitor is designed to measure exposure to chemicals in order to demonstrate workplace compliance with Permissible Exposure Limits (PELs) and Short Term Exposure Limits (STELs) defined by the Occupational Safety and Health Act of 1970 and Title 29 Code of Federal Regulations.

This **Technical Insert** contains product specific information on use and storage.

9159-575 11/09

Sampling Medium:	Activated molecular sieves, 5 Angstrom pore size
Analyte Sampled:	Nitrous Oxide
Analytical Method:	AT SOP L575: Desorption with water; headspace analysis by gas chromatography with electron capture detector (GC/ECD)
Recommended Sampling Time:	Sample during time the gas is in use Functional range: 15 minutes - 8 hours
Recommended Holding Time:	Monitors must be received by Lab within one (1) week after sampling.
Sample Capacity:	70 ppm for 8 hour sample (560 ppm-hrs)
Detection Limit (DL):	0.3 ppm for 8 hour sample (2.4 ppm-hrs)
Sampling Rate (SR):	0.75 mL/min
Interfering Substance(s):	Halogenated anesthetic gases, CFC's and HCFC's do not interfere when tested at their PELs.
Effect of Temperature:	Effect on result \leq 5% within 0 - 50 °C (32 - 122 °F).
Effect of Humidity:	Functions as claimed within 10 - 80% RH.
Accuracy (MTE):	Meets or exceeds NIOSH guidelines for accuracy: Maximum Total Error (MTE) \leq 25%.
Storage Conditions:	Store at controlled room temperature. Do Not Use after Expiration Date printed on pouch.
IMPORTANT:	Warranty valid only if Instructions for Use have been followed.
WARNING:	Wash affected area thoroughly if Sampling Medium is contacted.

For detailed directions, see **Instructions for Use** included in each package of Monitors.

For detailed information about the ChemDisk™ Personal Monitor, refer to www.assaytech.us or

call **TOLL FREE NUMBER 1-800-833-1258**