

TECHNICAL INSERT

Item No	587	ChemDisk™ Monitor for Hydrogen Peroxide	Prepared by	SCW
Edition	10/14/2010		Approved by	CRM

The ChemDisk™ Personal Monitor is designed to measure exposure to chemicals in order to demonstrate workplace compliance with Permissible Exposure Limits (PELs) and Short Term Exposure Limits (STELs) defined by the Occupational Safety and Health Act of 1970 and Title 29 Code of Federal Regulations

This **Technical Insert** contains product specific information on use and storage.

9159-587 10/10

Sampling Medium:	Chemically Activated Silica Gel Adsorbent
Analyte Sampled:	Hydrogen Peroxide
Analytical Method:	Similar to OSHA VI-6: Colorimetric analysis
Recommended Sampling Time:	PEL = 8 hrs or full shift Functional Range: 15 minutes – 8 hours
Recommended Holding Time:	Monitors must be received by Lab within two (2) weeks after sampling.
Sample Capacity:	10 ppm for 8 hour sample (80 ppm-hrs)
Detection Limit (DL):	0.02 ppm for 8 hour sample (0.16 ppm-hrs)
Sampling Rate (SR):	11.7 mL/min
Interfering Substance(s):	Compounds that form hydrogen peroxide in the presence of acids will interfere. Tests can be made for particular chemicals that may be suspected.
Effect of Temperature:	Effect on result \leq 5% within 0 - 50 °C (32 - 122 °F).
Effect of Humidity:	Functions as claimed from 30 - 92% RH
Accuracy (MTE):	Meets or exceeds OSHA requirements for accuracy: Maximum Total Error (MTE) \leq 19% at 1 ppm for 8 hours.
Storage Conditions:	Store at controlled room temperature. Do Not Use after Expiration Date printed on pouch.
IMPORTANT:	Warranty valid only if Instructions for Use have been followed.
WARNING:	Wash affected thoroughly area if Sampling Medium is contacted.

For detailed directions, see **Instructions for Use** included in each package of Monitors.

For detailed information about the ChemDisk™ Personal Monitor, refer to www.assaytech.us or

call **TOLL FREE NUMBER 1-800-833-1258**