

TECHNICAL INSERT

Item No	593	ChemDisk™ Monitor for	Prepared by	SCW
Edition	11/19/2009	Mercury Vapor	Approved by	CRM

The ChemDisk™ Personal Monitor is designed to measure exposure to chemicals in order to demonstrate workplace compliance with Permissible Exposure Limits (PELs) and Short Term Exposure Limits (STELs) defined by the Occupational Safety and Health Act of 1970 and Title 29 Code of Federal Regulations

This **Technical Insert** contains product specific information on use and storage.

9159-593 11/09

Sampling Medium:	Ultra high-purity gold
Analyte Sampled:	Mercury Vapor
Analytical Method:	Modified OSHA ID-140: Acid digestion; analysis of elemental mercury by cold vapor atomic fluorescence (CVAf).
Recommended Sampling Time:	PEL = 8 hrs or full shift Functional range: 1 – 12 hours
Recommended Holding Time:	Monitors must be received by Lab within four (4) weeks after sampling.
Sample Capacity:	15 mg/m ³ for 8 hour sample (120 mg/m ³ - hrs)
Detection Limit (DL):	0.001 mg/m ³ for 8 hours (0.008 mg/m ³ - hrs)
Sampling Rate (SR):	14.9 mL/min
Interfering Substance(s):	Monitor may absorb hydrogen sulfide and chlorine but these chemicals do not interfere in the analysis.
Effect of Temperature:	Effect on result ≤ 5% within 0 - 50 °C (32 - 122 °F)
Effect of Humidity:	Functions as claimed within 10 - 80% RH
Accuracy (MTE):	Meets or exceeds OSHA requirements for accuracy: Maximum Total Error (MTE) ≤ 10% at 0.05 mg/m ³ for 8 hours.
Storage Conditions:	Store at controlled room temperature. Do Not Use after Expiration Date printed on pouch. Best stored away from Mercury sources.
IMPORTANT:	Warranty valid only if Instructions for Use have been followed.
WARNING:	Wash affected area thoroughly if Sampling Medium is contacted.

For detailed directions, see **Instructions for Use** included in each package of Monitors.

For detailed information about the ChemDisk™ Personal Monitor, refer to www.assaytech.us or

call **TOLL FREE NUMBER 1-800-833-1258**