



AIHA Proficiency Analytical Testing Programs, LLC

2700 Prosperity Avenue, Suite 250, Fairfax, Virginia 22031 USA

main 1+ 703-846-0757 fax 1+ 703-207-8558

www.aihapat.org

2/5/2010

Steve Green
AT Labs, a Unit of Assay Technology
250 DeBartolo Place
Suite 2525
Boardman, OH 44512

Lab ID# 100903

Dear Steve Green,

Please find your laboratory's **Diffusive Sampler** results for the Industrial Hygiene Proficiency Analytical Testing (IHPAT) **Round 180**. It is the laboratory's responsibility to thoroughly review results and to immediately contact the AIHA Proficiency Analytical Testing Programs, LLC in writing, if any errors are found in your report.

The proficiency demonstrated by the results of this IHPAT Diffusive Sampler round is valid until the close of the retest round if the laboratory chooses to participate or the close of the next IHPAT Diffusive Sampler round on August 16, 2010. Unacceptable performance may be improved by correctly analyzing a set of retest samples. Retest Order Forms are available online at www.aihapat.org. The deadline to order a retest is February 26, 2010. Retest sample kits are mailed via USPS Mail within five days after the order deadline.

IHPAT Diffusive Sampler **Round 182** sample kits will be mailed to laboratories around July 1, 2010. Your laboratory's data will be due by 11:59pm EST on August 2, 2010.

Please handle, store and analyze your laboratory's PAT samples in the same manner as routine client samples. To submit results, visit the Proficiency Analytical Testing (PAT) page and click on the PAT Data Entry Portal: www.aihapat.org. **Always print and save the confirmation page** after submitting data via the PAT Data Entry Portal.

Your laboratory's password, needed to access the reports through the PAT Data Entry Portal is provided in the upper right hand corner (next to your lab ID#) of the address label on the results worksheet included with your PAT samples.

I encourage you to contact me with any feedback, questions or if you wish to contest your results at nmugambwa@aiha.org.

Sincerely,

A handwritten signature in black ink that reads 'Nmugambwa'. The signature is written in a cursive, flowing style.

Natasha Mugambwa, MS
Manager, AIHA PAT Programs, LLC

Industrial Hygiene Proficiency Analytical Testing Results

This document contains three sub-reports relating to Diffusive Round 180. The first report contains your laboratory's results listed per contaminant, per sample. The second report contains your current and 2 previous test rounds performance respectively (where applicable), and the final report contains summary results for all laboratories for Diffusive Round 180.

Testing Results for Diffusive Round 180

This part of the report contains your laboratory's results listed per contaminant, per sample.

Contaminant	Units	#	Result	Ref. Value	Lower Limit	Upper Limit	z-Score	Rating
Benzene (BNZ)	ppm	1	18.00	19.64	16.10	23.17	-1.4	A
	ppm	2	21.00	22.26	18.25	26.26	-0.9	A
O-xylene (OXY)	ppm	1	26.00	27.55	22.59	32.50	-0.9	A
	ppm	2	10.00	11.50	9.43	13.57	-2.2	A
Toluene (TOL)	ppm	1	40.00	41.18	33.77	48.60	-0.5	A
	ppm	2	14.00	15.00	12.30	17.70	-1.1	A

Please note:

Reference value is the mean of the reference laboratories.

Lower limit = reference value - 3 standard deviations

Upper limit = reference value + 3 standard deviations

Z-Score = (reported result - reference value)/standard deviation

A: Acceptable Analysis; U: Unacceptable Analysis

The acceptability of reported results is based on upper and lower performance limits.

Overall Performance Summary Concluding with 180

The following table contains your laboratory's current and 2 previous test rounds performance respectively (where applicable). For more information in regard to the determination of proficiency, please visit http://www.aiha.org/lc_pat/programs_ihpat.htm

Sample	Round	Round Performance	Round Score	Proficiency Status -Three Round Score
3M	176	6/6	Pass	
	178	6/6	Pass	
	180	6/6	Pass	P

Please note:

The denominators represent the total number of samples analyzed.

The numerators represent the number of acceptable results.

Pass: Round Score \geq 75% Fail: Round Score $<$ 75%

P – Proficient; NP – Non-proficient; I – Indeterminate.

A laboratory is rated proficient (P) for the associated FoT/Method(s), if the laboratory has a passing score for the applicable PT analyte class in two (2) of the last three (3) consecutive PT rounds. A laboratory is rated non-proficient (NP) for the applicable FoT/Method if the laboratory has failing scores for the associated PT analyte class in two (2) of the last three (3) consecutive PT rounds.

If a laboratory receives samples and does not report the data, the results will be treated as outliers

